


A-25	threo-3-Phenylserine	
		
<p>OA-6100 $\alpha = 2.20$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (95/5)] Flow rate : 1mL/min Detector : UV 254nm Injection : 5μL (5mg/mL)</p>	<p>OA-6000 $\alpha = 1.62$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (95/5)] Flow rate : 1mL/min Detector : UV 254nm</p>	
<p>OA-4800 $\alpha = 1.18$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/1,2-dichloroethane/ methanol/ trifluoroacetic acid (50/45/5/0.2) Flow rate : 1mL/min Detector : UV 254nm Injection : 3μL (4mg/mL)</p>	<p>OA-5000 $\alpha = 1.15$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/methanol (85/15)] Flow rate : 1mL/min Detector : UV 254nm</p>	
<p>OA-4900 $\alpha = 1.08$</p> <p>Column : 4mm i.d.×250mm Mobile phase : hexane/1,2-dichloroethane/ methanol/ trifluoroacetic acid (50/40/10/1) Flow rate : 1mL/min Detector : UV 254nm</p>		