



B-10	Atropine	
		
<p>OA-4500 $\alpha = 1.06$</p> <p>Column : 4mm i.d.×250mm×2 Mobile phase : hexane/1,2-dichloroethane/ methanol/trifluoroacetic acid (250/140/10/1) Flow rate : 1mL/min Detector : UV 254nm</p> 	<p>AGP $\alpha = 1.29$</p> <p>Column : 4mm i.d.×100mm Mobile phase : 2% 2-propanol in 10 mmol/L sodium phosphate buffer (pH 7.0) + 5mmol/L octanoic acid Flow rate : 0.9mL/min Detector : UV 265nm</p> 