





B-18	Normetanephrine	
		
<p><b>OA-6000</b> <span style="float: right;"><math>\alpha = 1.28</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            0.5mmol/L copper( II ) sulfate            in water            Flow rate : 1mL/min            Detector : UV 254nm</p> <div style="text-align: center;">  </div>	<p><b>OA-5000</b> <span style="float: right;"><math>\alpha = 1.15</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            0.2mmol/L copper( II ) acetate            and 0.1mol/L ammonium            acetate in water (pH 6.5)            Flow rate : 1mL/min            Detector : UV 280nm</p> <div style="text-align: center;">  </div>	
<p><b>CBH</b> <span style="float: right;"><math>\alpha = 2.05</math></span></p> <p>Column : 4mm i.d.×100mm            Mobile phase :            5% 2-propanol in 10mmol/L            sodium phosphate buffer            (pH 6.0) + 50 μmol/L EDTA            Flow rate : 0.9mL/min            Detector : UV 210nm</p> <div style="text-align: center;">  </div>		