


B-27	erythro-2-Amino-1,2-diphenylethanol	
		
<p>OA-7700 $\alpha = 1.80$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : methanol/20mmol/L potassium phosphate buffer (pH 3.0) (5/95) Flow rate : 0.5mL/min Detector : UV 210nm Injection : 1μL (2mg/mL)</p>	<p>OA-7000 $\alpha = 1.20$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : methanol/20mmol/L potassium phosphate buffer (pH 3.0) (5/95) Flow rate : 0.5mL/min Detector : UV 254nm Injection : 2μL (10mg/mL)</p>	
<p>OA-4900 $\alpha = 1.20$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/trifluoroacetic acid (100/10/5/0.1) Flow rate : 1mL/min Detector : UV 254nm</p>	<p>OA-4500 $\alpha = 1.14$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/trifluoroacetic acid (100/10/5/0.1) Flow rate : 1mL/min Detector : UV 254nm</p>	
