


G-02	Valine	
		
<p><b>OA-6100</b> <span style="float: right;"><math>\alpha=7.55</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            2mmol/L copper( II ) sulfate            in [water/acetonitrile (98/2)]            Flow rate : 1mL/min            Detector : UV 254nm            Injection : 3<math>\mu</math>L (3mg/mL)</p>	<p><b>OA-6000</b> <span style="float: right;"><math>\alpha=4.64</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            2mmol/L copper( II ) sulfate            in [water/acetonitrile (95/5)]            Flow rate : 1mL/min            Detector : UV 254nm</p>	
<p><b>OA-5000</b> <span style="float: right;"><math>\alpha=1.71</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            2mmol/L copper( II ) sulfate            in [water/2-propanol (95/5)]            Flow rate : 1mL/min            Detector : UV 254nm            Injection : 3<math>\mu</math>L (5mg/mL)</p>		

