


H-11

Tetrahydro-2-furancarboxylic acid

**OA-6000** $\alpha=1.70$


Column : 4.6mm i.d.×150mm

Mobile phase :

2mmol/L copper(II) sulfate
in [water/acetonitrile (90/10)]

Flow rate : 1mL/min

Detector : UV 254nm

Injection : 3 μ L (6mg/mL)**OA-6100** $\alpha=1.28$


Column : 4.6mm i.d.×150mm

Mobile phase :

2mmol/L copper(II) sulfate
in [water/acetonitrile (95/5)]

Flow rate : 1mL/min

Detector : UV 254nm

Injection : 3 μ L (5mg/mL)**OA-5000** $\alpha=1.06$

Column : 4.6mm i.d.×150mm

Mobile phase :

1mmol/L copper(II) sulfate
in water

Flow rate : 1mL/min

Detector : UV 254nm

Injection : 3 μ L (5mg/mL)