



H-12	Tetrahydro-3-furancarboxylic acid	
		
<p>OA-6000 $\alpha=1.10$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 1mmol/L copper(II) sulfate in water Flow rate : 1mL/min Detector : UV 254nm Injection : 3μL (8mg/mL)</p> 	<p>OA-5000 $\alpha=1.05$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 1mmol/L copper(II) sulfate in water Flow rate : 1mL/min Detector : UV 254nm Injection : 3μL (7mg/mL)</p> 