





C-01	1-Phenylethylamine	
		
<p>OA-7700 $\alpha=1.20$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : methanol/20mmol/L potassium phosphate buffer (pH 3.0) (2/98) Flow rate : 0.5mL/min Detector : UV 210nm Injection : 1μL (2mg/mL)</p> 	<p>OA-4100 $\alpha=1.08$</p> <p>Column : 4mm i.d.×250mm×2 Mobile phase : hexane/ethanol/ trifluoroacetic acid (240/10/1) Flow rate : 1mL/min Detector : UV 254nm</p> 	
<p>OA-4700 $\alpha=1.08$</p> <p>Column : 4mm i.d.×250mm Mobile phase : hexane/ethanol/ trifluoroacetic acid (240/10/1) Flow rate : 1mL/min Detector : UV 254nm</p> 	<p>OA-4900 $\alpha=1.06$</p> <p>Column : 4mm i.d.×250mm Mobile phase : hexane/ethanol/ trifluoroacetic acid (240/10/1) Flow rate : 1mL/min Detector : UV 254nm</p> 