



H-10	2-Hydroxyhexanoic acid	
		
<p>OA-6100 $\alpha=2.43$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (90/10)] Flow rate : 1mL/min Detector : UV 254nm Injection : 1μL (6mg/mL)</p> 	<p>OA-5000 $\alpha=1.53$</p> <p>Column : 4.6mm i.d.×50mm Mobile phase : 2mmol/L copper(II) sulfate in [water/2-propanol (85/15)] Flow rate : 1mL/min Detector : UV 254nm Injection : 5μL (6mg/mL)</p> 