





A-20	N-Benzoylvaline	
		
<p>OA-3300 $\alpha = 1.19$</p> <p>Column : 4mm i.d.×250mm Mobile phase : 0.01mol/L ammonium acetate in methanol Flow rate : 1mL/min Detector : UV 254nm</p> 	<p>OA-3100 $\alpha = 1.12$</p> <p>Column : 4mm i.d.×250mm Mobile phase : 0.005mol/L ammonium acetate in methanol Flow rate : 1mL/min Detector : UV 254nm</p> 	
<p>OA-4600 $\alpha = 1.07$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/ trifluoroacetic acid (90/5/5/0.2) Flow rate : 1mL/min Detector : UV 254nm</p> 	<p>AGP $\alpha = 2.16$</p> <p>Column : 4mm i.d.×100mm Mobile phase : 2-propanol/ 10mmol/L sodium phosphate buffer (pH 5.0) (3/97) Flow rate : 0.9mL/min Detector : UV 200nm</p> 