










A-21	N-Benzoyl-phenylalanine	
		
<p>OA-3300 $\alpha=1.17$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : 0.01mol/L ammonium acetate in methanol Flow rate : 1mL/min Detector : UV 254nm Injection : 0.4µL (10mg/mL)</p> 	<p>OA-4700 $\alpha=1.16$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/ trifluoroacetic acid (90/5/5/0.2) Flow rate : 1mL/min Detector : UV 254nm</p> 	
<p>OA-4600 $\alpha=1.10$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/ trifluoroacetic acid (90/5/5/0.2) Flow rate : 1mL/min Detector : UV 254nm Injection : 1µL (2mg/mL)</p> 	<p>OA-4100 $\alpha=1.07$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/ trifluoroacetic acid (90/5/5/0.2) Flow rate : 1mL/min Detector : UV 254nm Injection : 0.5µL (6mg/mL)</p> 	
<p>OA-3100 $\alpha=1.07$</p> <p>Column : 4mm i.d.×250mm Mobile phase : 0.005mol/L ammonium acetate in methanol Flow rate : 1mL/min Detector : UV 254nm</p> 	<p>OA-3200 $\alpha=1.06$</p> <p>Column : 4mm i.d.×250mm Mobile phase : 0.01mol/L ammonium acetate in methanol Flow rate : 1mL/min Detector : UV 254nm</p> 	

A-21	N-Benzoyl-phenylalanine	
 <chem>O=C(NCc1ccccc1)C(=O)c2ccccc2</chem>		
<p>OA-4500 $\alpha = 1.04$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/ trifluoroacetic acid (90/5/5/0.2) Flow rate : 1mL/min Detector : UV 254nm</p> 	<p>OA-4400 $\alpha = 1.04$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/ trifluoroacetic acid (90/5/5/0.2) Flow rate : 1mL/min Detector : UV 254nm</p> 