




B-05		Oxprenolol	
			
<p>OA-4100</p> <p>Column : 4.6mm i.d.×250mm</p> <p>Mobile phase : hexane/tetrahydrofuran/ methanol/trifluoroacetic acid (60/38/2/0.2)</p> <p>Flow rate : 1mL/min</p> <p>Detector : UV 254nm</p> <p>Injection : 5µL (1mg/mL)</p>	<p>$\alpha = 1.09$</p>	<p>OA-4700</p> <p>Column : 4.6mm i.d.×250mm</p> <p>Mobile phase : hexane/tetrahydrofuran/ methanol/trifluoroacetic acid (60/38/2/0.2)</p> <p>Flow rate : 1mL/min</p> <p>Detector : UV 254nm</p> <p>Injection : 5µL (1mg/mL)</p>	<p>$\alpha = 1.09$</p>
			
<p>AGP</p> <p>Column : 4mm i.d.×100mm</p> <p>Mobile phase : 2-propanol/10mmol/L sodium acetate buffer (pH 4.5) (1/99) (total acetate conc.=25mmol/L)</p> <p>Flow rate : 0.9mL/min</p> <p>Detector : UV 225nm</p>	<p>$\alpha = 1.29$</p>		
