


B-17	Norphenylephrine	
		
<p>OA-7700 $\alpha = 1.43$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : 20mmol/L KH₂PO₄ Flow rate : 0.5mL/min Detector : UV 254nm Injection : 2μL (2mg/mL)</p>	<p>OA-5000 $\alpha = 1.31$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 0.2mmol/L copper(II) acetate and 0.1mol/L ammonium acetate in water (pH 6.5) Flow rate : 1mL/min Detector : UV 280nm</p>	
<p>OA-6000 $\alpha = 1.20$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 0.5mmol/L copper(II) sulfate in water Flow rate : 1mL/min Detector : UV 254nm</p>		

