


B-23	metoprolol	
		
<p><b>OA-4900</b> <span style="float: right;"><math>\alpha = 1.09</math></span></p> <p>Column : 4mm i.d.×250mm            Mobile phase :            hexane/1,2-dichloroethane/            ethanol/trifluoroacetic acid            (240/140/20/1)            Flow rate : 1mL/min            Detector : UV 254nm</p>	<p><b>AGP</b> <span style="float: right;"><math>\alpha = 1.26</math></span></p> <p>Column : 4mm i.d.×100mm            Mobile phase :            2-propanol/10mmol/L sodium            phosphate buffer (pH 7.0) (0.5/99.5)            Flow rate : 0.9mL/min            Detector : UV 225nm</p>	
<p><b>CBH</b> <span style="float: right;"><math>\alpha = 1.26</math></span></p> <p>Column : 4mm i.d.×100mm            Mobile phase :            5% 2-propanol in 10mmol/L            sodium phosphate buffer            (pH 6.0) + 50 <math>\mu</math>mol/L EDTA            Flow rate : 0.9mL/min            Detector : UV 225nm</p>		

