


G-08

Glutamine

**OA-6100** $\alpha=3.04$ 


Column : 4.6mm i.d.×150mm

Mobile phase :

2mmol/L copper( II ) sulfate  
in [water/acetonitrile (98/2)]

Flow rate : 1mL/min

Detector : UV 254nm

Injection : 2 $\mu$ L (4mg/mL)**OA-6000** $\alpha=2.81$ 


Column : 4.6mm i.d.×150mm

Mobile phase :

1mmol/L copper( II ) sulfate  
in water

Flow rate : 1mL/min

Detector : UV 254nm

**OA-5000** $\alpha=1.71$ 

Column : 4.6mm i.d.×150mm

Mobile phase :

1mmol/L copper( II ) sulfate  
in water

Flow rate : 1mL/min

Detector : UV 254nm

