


G-17	Threonine	
		
<p><b>OA-6100</b> <span style="float: right;"><math>\alpha=2.35</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            1mmol/L copper( II ) sulfate            in water            Flow rate : 1mL/min            Detector : UV 254nm            Injection : 3<math>\mu</math>L (3mg/mL)</p>	<p><b>OA-6000</b> <span style="float: right;"><math>\alpha=1.94</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            1mmol/L copper( II ) sulfate            in water            Flow rate : 1mL/min            Detector : UV 254nm</p>	<p><b>OA-5000</b> <span style="float: right;"><math>\alpha=1.20</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            0.5mmol/L copper( II ) acetate            in water            Flow rate : 0.5mL/min            Detector : UV 254nm</p>
