


G-20

Asparagine

**OA-6000** $\alpha=1.49$


Column : 4.6mm i.d.×150mm

Mobile phase :

1mmol/L copper(II) sulfate
in water

Flow rate : 1mL/min

Detector : UV 254nm

Injection : 1.5 μ L (3mg/mL)**OA-6100** $\alpha=1.43$


Column : 4.6mm i.d.×150mm

Mobile phase :

1mmol/L copper(II) sulfate
in water

Flow rate : 1mL/min

Detector : UV 254nm

Injection : 3 μ L (3mg/mL)**OA-5000** $\alpha=1.10$

Column : 4.6mm i.d.×150mm, 60°C

Mobile phase :

1mmol/L copper(II) sulfate
in water

Flow rate : 0.5mL/min

Detector : UV 254nm

