



G-22		N-Acetylalanine	
			
<b>OA-4700</b>	$\alpha=1.17$	<b>OA-5000</b>	$\alpha=1.17$
Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/ trifluoroacetic acid (90/5/5/0.2) Flow rate : 1mL/min Detector : UV 254nm Injection : 3μL (12mg/mL)		Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper( II ) sulfate in [water/acetonitrile (95/5)] Flow rate : 1mL/min Detector : UV 254nm	