


A-04	Histidine	
		
<p>OA-6100 $\alpha=2.52$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (95/5)] Flow rate : 1mL/min Detector : UV 254nm Injection : 3μL (6mg/mL)</p>	<p>OA-6000 $\alpha=2.06$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (85/15)] Flow rate : 1mL/min Detector : UV 254nm</p>	
<p>OA-5000 $\alpha=1.37$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/2-propanol (95/5)] Flow rate : 1mL/min Detector : UV 254nm Injection : 3μL (6mg/mL)</p>	<p>OA-8000 $\alpha=1.14$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/ trifluoroacetic acid/water (80/15/5/0.5/0.5) Flow rate : 0.7mL/min Detector : UV 220nm</p>	
