


E-15

Hydrobenzoin

**OA-2500I** $\alpha=1.13$


Column : 4.6mm i.d.×250mm

Mobile phase :

hexane/ethanol (95/5)

Flow rate : 1mL/min

Detector : UV 254nm

Injection : 10 μ L (1mg/mL)**OA-7000** $\alpha=1.28$


Column : 4.6mm i.d.×250mm

Mobile phase:

methanol/water (30/70)

Flow rate : 0.5mL/min

Detector : UV 254nm

Injection : 5 μ L (1mg/mL)**OA-7700** $\alpha=1.21$

Column : 4.6mm i.d.×250mm

Mobile phase:

methanol/water (50/50)

Flow rate : 0.5mL/min

Detector : UV 254nm

Injection : 5 μ L (10mg/mL)