


G-12	Pipecolic acid	
		
<p><b>OA-5000</b> <span style="float: right;"><math>\alpha=1.77</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            1mmol/L copper( II ) sulfate            in water            Flow rate : 1mL/min            Detector : UV 254nm</p>	<p><b>OA-6000</b> <span style="float: right;"><math>\alpha=1.42</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            1mmol/L copper( II ) sulfate            in water            Flow rate : 1mL/min            Detector : UV 254nm</p>	<p><b>OA-6100</b> <span style="float: right;"><math>\alpha=1.14</math></span></p> <p>Column : 4.6mm i.d.×150mm            Mobile phase :            1mmol/L copper( II ) sulfate            in water            Flow rate : 1mL/min            Detector : UV 254nm            Injection : 5<math>\mu</math>L (1.5mg/mL)</p>
