



G-47	3-Aminopyrrolidine	
		
<p>OA-6000 $\alpha=2.61$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 1mmol/L copper(II) sulfate in water Flow rate : 1mL/min Detector : UV 254nm Injection : 3μL (6mg/mL)</p> 	<p>OA-6100 $\alpha=1.74$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 1mmol/L copper(II) sulfate in water Flow rate : 1mL/min Detector : UV 254nm Injection : 3μL (6mg/mL)</p> 