


H-16

Pantoic acid

**OA-5000** $\alpha = 1.63$

Column : 4.6mm i.d.×150mm

Mobile phase :

2mmol/L copper(II) sulfate
in [water/2-propanol (90/10)]

Flow rate : 1mL/min

Detector : UV 254nm

Injection : 5 μ L (8mg/mL)