


A-05	Phenylglycine		
			
<p>OA-6100 $\alpha = 3.46$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (90/10)] Flow rate : 1mL/min Detector : UV 254nm Injection : 1μL (1.5mg/mL)</p>	<p>OA-6000 $\alpha = 1.89$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (85/15)] Flow rate : 1mL/min Detector : UV 254nm</p>	<p>OA-5000 $\alpha = 1.63$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/2-propanol (95/5)] Flow rate : 1mL/min Detector : UV 254nm Injection : 5μL (1.5mg/mL)</p>	<p>OA-4800 $\alpha = 1.16$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/1,2-dichloroethane/ methanol/trifluoroacetic acid (50/45/5/0.2) Flow rate : 1mL/min Detector : UV 254nm Injection : 3μL (4mg/mL)</p>
