




B-07	Propafenone	
		
<p>OA-4800 $\alpha = 1.12$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/tetrahydrofuran/ methanol/trifluoroacetic acid (60/35/5/0.2) Flow rate : 1mL/min Detector : UV 254nm Injection : 2µL (1mg/mL)</p> 	<p>AGP $\alpha = 1.12$</p> <p>Column : 4mm i.d.×100mm Mobile phase : 1-propanol/10mmol/L sodium phosphate buffer (pH 6.0) (8/92) Flow rate : 0.5mL/min Detector : UV 254nm Injection : 2µL (1mg/mL)</p> 	
<p>CBH $\alpha = 2.03$</p> <p>Column : 4mm i.d.×100mm Mobile phase : 5% 2-propanol in 10mmol/L sodium phosphate buffer (pH 7.0) + 50µmol/L EDTA Flow rate : 0.9mL/min Detector : UV 225nm</p> 