



B-21	Phenylglycinol	
		
<p>OA-7700</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : 20mmol/L potassium phosphate buffer (pH 3.0) Flow rate : 0.5mL/min Detector : UV 254nm Injection : 2μL (10mg/mL)</p>	<p>OA-8000</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : acetonitrile/perchloric acid aq. (pH 1.5) (10/90) Flow rate : 0.7mL/min Detector : UV 254nm Injection : 2μL (10mg/mL)</p>	
<p>$\alpha = 1.29$</p> 	<p>$\alpha = 1.08$</p> 