




B-28	Synephrine	
		
<p>OA-7700 $\alpha=1.20$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : 20mmol/L KH₂PO₄ Flow rate : 0.5mL/min Detector : UV 230nm Injection : 1μL (2mg/mL)</p> 	<p>OA-5000 $\alpha=1.19$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : 1mmol/L copper(II) acetate and 20mmol/L ammonium acetate in water (pH 6.5) Flow rate : 0.8mL/min Detector : UV 254nm Injection : 3μL (4mg/mL)</p> 	
<p>OA-4900 $\alpha=1.10$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/1,2-dichloroethane/ methanol/trifluoroacetic acid (240/140/20/1) Flow rate : 1mL/min Detector : UV 254nm Injection : 1μL (5mg/mL)</p> 