


C-05	Chlorpheniramine	
		
<p>OA-4500 $\alpha = 1.21$</p> <p>Column : 4mm i.d.×250mm Mobile phase : hexane/ethanol/ trifluoroacetic acid (100/60/0.6) Flow rate : 1mL/min Detector : UV 254nm</p>	<p>OA-7000 $\alpha = 1.12$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : acetonitrile/20mmol/L potassium phosphate buffer (pH 3.5) (5/95) Flow rate : 0.5mL/min Detector : UV 254nm Injection : 1µL (2mg/mL)</p>	
<p>OA-4100 $\alpha = 1.12$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/methanol/ trifluoroacetic acid (80/15/5/0.2) Flow rate : 1mL/min Detector : UV 254nm Injection : 5µL (1mg/mL)</p>		

