



C-21	PTH-tryptophan	
		
<p><b>OA-4700</b> <math>\alpha = 1.33</math></p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/ methanol/trifluoroacetic acid (80/15/5/0.2) Flow rate : 1mL/min Detector : UV 254nm Injection : 1<math>\mu</math>L (1mg/mL)</p> 	<p><b>OA-4100</b> <math>\alpha = 1.26</math></p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/2-propanol/ methanol/trifluoroacetic acid (80/15/5/0.2) Flow rate : 1mL/min Detector : UV 254nm Injection : 1<math>\mu</math>L (1mg/mL)</p> 