


D-09	Mandelic acid	
		
<p>OA-6100 $\alpha=2.15$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (90/10)] Flow rate : 1mL/min Detector : UV 254nm Injection : 2μL (16mg/mL)</p>	<p>OA-5000 $\alpha=1.24$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (85/15)] Flow rate : 1mL/min Detector : UV 254nm</p>	<p>OA-4500 $\alpha=1.05$</p> <p>Column : 4.6mm i.d.×250mm Mobile phase : hexane/1,2-dichloroethane/ methanol/trifluoroacetic acid (80/18/2/0.2) Flow rate : 1mL/min Detector : UV 254nm Injection : 5μL (7mg/mL)</p>
