


G-01

Alanine

**OA-6000** $\alpha=2.28$


Column : 4.6mm i.d.×150mm

Mobile phase :

1mmol/L copper(II) sulfate
in water

Flow rate : 1mL/min

Detector : UV 254nm

**OA-6100** $\alpha=2.23$


Column : 4.6mm i.d.×150mm

Mobile phase :

1mmol/L copper(II) sulfate
in water

Flow rate : 1mL/min

Detector : UV 254nm

Injection : 5 μ L (2mg/mL)**OA-5000** $\alpha=1.58$

Column : 4.6mm i.d.×150mm

Mobile phase :

1mmol/L copper(II) sulfate
in water

Flow rate : 1mL/min

Detector : UV 254nm

