


G-05	Serine	
		
<p>OA-6100 $\alpha=1.82$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 1mmol/L copper(II) sulfate in water Flow rate : 1mL/min Detector : UV 254nm Injection : 5μL (2mg/mL)</p>	<p>OA-6000 $\alpha=1.70$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 1mmol/L copper(II) sulfate in water Flow rate : 1mL/min Detector : UV 254nm</p>	
<p>OA-5000 $\alpha=1.17$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 1mmol/L copper(II) sulfate in water Flow rate : 0.2mL/min Detector : UV 254nm</p>		

