


G-10	Arginine	
		
<p>OA-6100 $\alpha=6.50$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (98/2)] Flow rate : 1mL/min Detector : UV 254nm Injection : 1μL (5mg/mL)</p>	<p>OA-6000 $\alpha=2.39$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 2mmol/L copper(II) sulfate in [water/acetonitrile (95/5)] Flow rate : 1mL/min Detector : UV 254nm</p>	
<p>OA-5000 $\alpha=2.15$</p> <p>Column : 4.6mm i.d.×150mm Mobile phase : 1mmol/L copper(II) sulfate in water Flow rate : 1mL/min Detector : UV 254nm</p>		

