


## G-51 Pantothenic acid

**OA-6100** $\alpha = 1.12$ 

Column : 4.6mm i.d.×150mm×2

Mobile phase :

2mmol/L copper( II ) sulfate

in [water/acetonitrile (95/5)]

Flow rate : 1mL/min

Detector : UV 254nm

