


I-07

Propargyllone

**OA-4100** $\alpha=1.06$


Column : 4.6mm i.d.×250mm

Mobile phase :

hexane/ethanol (98/2)

Flow rate : 1mL/min

Detector : UV 230nm

Injection : 1 μ L (1mg/mL)**OA-4500** $\alpha=1.06$


Column : 4.6mm i.d.×250mm

Mobile phase :

hexane/ethanol (98/2)

Flow rate : 1mL/min

Detector : UV 230nm

Injection : 1 μ L (1mg/mL)**OA-4700** $\alpha=1.06$


Column : 4.6mm i.d.×250mm

Mobile phase :

hexane/ethanol (98/2)

Flow rate : 1mL/min

Detector : UV 230nm

Injection : 1 μ L (1mg/mL)**OA-7700** $\alpha=1.12$

Column : 4.6mm i.d.×250mm

Mobile phase : water

Flow rate : 0.5mL/min

Detector : UV 230nm

